

K110536

MAR 19 2012

510 (K) SUMMARY

This summary of 510(K) safety and effectiveness information is being submitted in accordance with requirements of SMDA 1990 and 21CFR 807.92.

1. Submitter's Identification:

Zibo Tengyue Plastic Products Co., Ltd.
77 Center of Zhangdian, Zhangdian District
Zibo, Shandong, 255400

Contact Person:

Jiang, Xiyan Product Manager
Phone Number: 011-86-13355336122

Date summary prepared: August 18, 2011

2. Name of the Device:

Zibo Tengyue Plastic Products Co., Ltd.
Powder-Free Vinyl Patient Examination Gloves, Non-Sterile

3. Predicate Device Information:

Shijiazhuang Star Plastic Co., Ltd.
Vinyl Examination Gloves – Powder Free (K100699)

Tangshan Zhonghong Pulin Food Products Co., Ltd.
Vinyl Examination Gloves – Powder Free (K022091)

4. Device Description:

A Powder-Free Vinyl Patient Examination Gloves is a disposable device intended for medical purposes that is worn upon the examiner's hands or finger to prevent contamination between patient and examiner (21CFR 880.6250). Our Powder-Free Vinyl Patient Examination Gloves do not contain any UPS powder according to the results of testing conducted on the gloves. The residual powder testing is conducted per standards of ASTM D-6124-06, which demonstrates that the powder free glove is less than 2mg/pc. Classified by FDA's General and Plastic Surgery Device panel as Class I, 21 CFR 880.6250, Powder-Free Vinyl Patient Examination Glove, 80LYZ, and meets all requirement of ASTM Standard D5250-06 (2011).

5. **Intended Use:**

A patient examination glove is a disposable device intended for medical purposes that is worn upon the examiner's hands or finger to prevent contamination between patient and examiner (21CFR 880.6250)

6. **Comparison to Predicate Devices:**

Zibo Tengyue Plastic Products Co., Ltd. Powder-Free Vinyl Patient Examination Gloves are substantially equivalent in safety and effectiveness to Tangshan Zhonghong Pulin Food Products Co., Ltd. (K022091) and Shijiazhuang Star Plastic, Co., Ltd. (K100699) Powder-Free Vinyl Patient Examination Gloves.

7. **Discussion of Non-Clinical tests performed for Determination of Substantial Equivalence are as follows:**

The standards used for Zibo Tengyue Plastic Products Co., Ltd. glove production are based on ASTM-D-5250-06 (2011). All testing meets requirements for Physical and Dimensions Testing conducted on gloves, Inspection Level S-2, AOL 2.5.

The FDA 1000 ml. Water Fill Test was also conducted with samplings of AOL 2.5, Inspection Level I, meeting these requirements, Primary Skin irritation and Skin Sensitization (allergic contact dermatitis) testing was conducted with results showing no primary skin irritant or sensitization reactions.

Our gloves meet standard of ASTM D6124-06 (Standard Test Method for Residual Powder on Medical Gloves).

8. **Discussion of Clinical Tests Performed:**

Not Applicable – There is no hypoallergenic claim.

9. **Conclusions:**

Zibo Tengyue Plastic Products Co., Ltd. Powder-Free Vinyl Patient Examination Gloves conforms fully to ASTM-D-5250-06 (2011) standard as well as applicable 21 CFR references, and, meets pinhole FDA requirements, biocompatibility requirements and labeling claims as shown by data in Section 7. There are no safety/efficacy issues or new claims from the "substantial equivalence" products cited.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room -WO66-G609
Silver Spring, MD 20993-0002

Zibo Tengyue Plastic Products Company, Limited
C/O Mr. Frank Liu
President
Basic Medical Industries, Incorporated
12390 East End Avenue
Chino, California 91710

MAR 19 2012

Re: K110536

Trade/Device Name: Powder-Free Vinyl Patient Examination Gloves, Non-Sterile
Regulation Number: 21 CFR 880.6250
Regulation Name: Patient Examination Glove
Regulatory Class: I
Product Code: LYZ
Dated: February 10, 2012
Received: February 24, 2012

Dear Mr. Liu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Anthony D. Watson, B.S., M.S., M.B.A.
Director

Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

INDICATIONS FOR USE

Applicant: Zibo Tengyue Plastic Products Co., Ltd.

510(k) Number: K110536

Device Name: Powder-Free Vinyl Patient Examination Gloves, Non-Sterile

Indications of Use:

A patient examination glove is a disposable device intended for medical purposes that is worn upon the examiner's hands or finger to prevent contamination between patient and examiner (21CFR 880.6250)

Prescription Use _____
(Part 21 CFR 801 Subpart D)

AND/OR

Over the Counter Use X
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

E. L. H. J. Chamberlain
(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K110536